

# SCHNEIDER LABORATORIES

INCORPORATED

2512 W. Cary Street • Richmond, Virginia • 23220-5117  
804-353-6778 • 800-785-LABS (5227) • (FAX) 804-359-1475

*Excellence in Service and Technology*

AIHA/ELLAP 100527, NVLAP 101150-0, NYELAP/NELAC 11413, CAELAP 2078, NC 593, SC 93003

## LABORATORY ANALYSIS REPORT

Asbestos and Other Fibers Counted based on NIOSH 7400 Method, Issue 2, Aug. 12, 1994

Using SLI A8

ACCOUNT #:  
CLIENT:  
ADDRESS:

DATE COLLECTED: 8/13/2008  
DATE RECEIVED: 8/19/2008  
DATE ANALYZED: 8/26/2008  
DATE REPORTED: 8/26/2008  
RESPIRATOR: Half Face

PROJECT NAME:  
JOB LOCATION:  
PROJECT NO.:  
PO NO.:

Area and Environmental Samples

SLI Sample No.	Client Sample No.	Sample Identification	Sample Date	Flow Rate (L/min)	Sample Time (min)	Sample Volume (L)	Fiber Count (f/field)	Actual Exposure (f/cc)
29860302	3	Outside Area	8/13/2008	2.00	150	300.0	< 0.055	< 0.009
29860303	4	Blank	8/13/2008				< 0.055	

Analyst:

Total Number of Pages in Report: 1

Results relate only to samples as received by the laboratory.

Reviewed By:

Visit [www.slabinc.com](http://www.slabinc.com) for current certifications.

OSHA PEL is 0.1 f/cc for 8 hour TWA. Est. limit of detection: 7 f/mm<sup>2</sup> or 0.055 f/field. Data precision justifies 2 sig. figures. For f/mm<sup>2</sup>, divide f/field by microscopic field area of 0.00785 mm<sup>2</sup>. Results are not blank-corrected unless noted by analyst. Exposure calculations are based on client-supplied information and assume zero exposure for time not sampled. Est. relative standard deviations: Intra-Laboratory: ± 0.30; Inter-Laboratory: ± 0.38. Unusual sample conditions, if any, are described.