

SCHNEIDER LABORATORIES

INCORPORATED

2512 W. Cary Street • Richmond, Virginia • 23220-5117
804-353-6778 • 800-785-LABS (5227) • (FAX) 804-353-6928

Excellence in Service and Technology

AIHA/ELLAP 100527, NVLAP 101150-0, NYELAP/NELAC 11413, CAELAP 2078, NC 593

LABORATORY ANALYSIS REPORT

Asbestos and Other Fibers Counted By NIOSH 7400 Method, Issue 2, Aug. 12, 1994

ACCOUNT: 1-03-10382
CLIENT: SLI
ADDRESS: 2512 West Cary Street
RICHMOND, VA 23225

DATE COLLECTED: 9/19/2003
DATE RECEIVED: 9/25/2003
DATE ANALYZED: 9/29/2003
DATE REPORTED: 10/15/2003
RESPIRATOR: H/F

PO NO.:
PROJECT NAME:
PROJECT NO.:
JOB LOCATION:

Area and Environmental Samples

SLI Sample No.	Client Sample No.	Sample Identification	Sample Date	Flow Rate (L/min)	Sample Time (min)	Sample Volume (L)	Fiber Count (f/field)	Actual Exposure (f/cc)
26923190	254	Sample 1	9/19/2003	2.00	72	144.0	< 0.055	< 0.019
26923193	257	Blank	9/19/2003	0.00	0	0.0	< 0.055	
26923194	258	Blank	9/19/2003	0.00	0	0.0	< 0.055	

ANALYST: HIND ELDANAF
Total no. of pages in report =

REVIEWED BY

SAMI A. HOSN

OSHA PEL is 0.1 f/cc for 8h TWA. Method limit: 0.01 f/cc. Results are not blank-corrected unless noted by analyst.
Exposure calculations are based on client-supplied information. 8 hour TWAs assume zero exposure for time not sampled.
Microscopic field area (mm²): 0.00785. Estimated limit of detection: 7 f/mm².
Estimated relative standard deviations: Intra-Laboratory: ± 0.26; Inter-Laboratory: ± 0.41.