



## **SCHNEIDER LABORATORIES INCORPORATED**

**Mailing & Shipping:** 2512 W. Cary Street, Richmond, VA 23220-5117

**Remit to:** P O Box 35702, Richmond, VA 23235

804-353-6778 Toll Free - 800-785-LABS FAX - 804-359-1475

www.slabinc.com [info@slabinc.com](mailto:info@slabinc.com)

**Excellence in Service and Technology Since 1987**

**Best Service, Best Quality, Best Price –**

**Guaranteed!**

*Woman Business Enterprise*

### **OPERATING INSTRUCTIONS GilAir5**

### **Constant High Flow Tri-Mode Air Sampler**

#### **General Information:**

Pump should be properly prepared which includes the charging and calibrating of the unit for proper flow rate. (Verify information on your Certificate of Calibration)

#### **STANDARD FEATURES:**

- Flow Rate 850cc – 5000cc
- Built-in Rotameter
- Rechargeable Battery
- Will Operate 10 hours @ 3.0 LPM
- Once the pump is turned off, the run data is displayed for 5 minutes, then shuts off.

#### **Operating Instructions:**

- Verify that the unit is properly charged
- Using a small Phillips screwdriver (provided with pump) back out the holding screw just enough so the anti-tamper cover plate can be rotated 180°. This exposes the ON/OFF switch and the flow adjusting screw. (See Certificate of Calibration for flow rate)
- Assemble the cassette filter holder for the sampling media. Use the appropriate filter for the sampling method.
- Move the ON/OFF switch to the ON position.
- Using the built-in rotameter on the BDX II as a flow indicator only, verify that the unit is flowing at the required rate.
- The unit is now ready for sample collection.
- Attach the unit to the individual's belt or pocket. Attach the cassette to the clip assembly and position it in the individual's breathing zone.
- Run sample for 8 hours or time designated by the collector. Document flow rate and sampling time on your chain-of-custody.