


# DIVISION OF WATER RESOURCES LABORATORY CERTIFICATION BRANCH

*In accordance with the provisions of N.C.G.S. 143-215.3 (a) (1), 143-215.3 (a)(10) and NCAC 2H.0800:*



# 2020

## Schneider Laboratories Global Inc.

*Is hereby certified to perform environmental analysis as listed on Attachment I and report monitoring data to DEQ for compliance with NPDES effluent, surface water, groundwater, and pretreatment regulations.*

*By reference 15A NCAC 2H .0800 is made a part of this certificate.*

*This certificate does not guarantee validity of data generated, but indicates the methodology, equipment, quality control procedures, records, and proficiency of the laboratory have been examined and found to be acceptable.*

This certificate shall be valid until 12/31/2020

Certificate No. 593

A handwritten signature in brown ink, appearing to read "Dana B. Satterwhite", is written over a horizontal line.

Dana B. Satterwhite

# North Carolina Wastewater/Groundwater Laboratory Certification

## Certified Parameters Listing

Lab Name: Schneider Laboratories Global Inc.  
Address: 2512 W. Cary Street  
Richmond, VA 23220-

Certificate Number: 593  
Effective Date: 1/1/2020  
Expiration Date: 12/31/2020  
Date of Last Amendment: 10/4/2018

The above named laboratory, having duly met the requirements of 15A NCAC 2H.0800, is hereby certified for the measurement of the parameters listed below.

### CERTIFIED PARAMETERS

#### METAL

##### ALUMINUM

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
SW-846 6010 D (Aqueous)

##### ANTIMONY

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)  
EPA 200.9, Rev. 2.2, 1994 (Aqueous)

##### ARSENIC

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)  
EPA 200.9, Rev. 2.2, 1994 (Aqueous)

##### BARIUM

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)

##### BERYLLIUM

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
SW-846 6010 D (Aqueous)

##### CADMIUM

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)  
EPA 200.9, Rev. 2.2, 1994 (Aqueous)

##### CHROMIUM TOTAL

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)

#### COBALT

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
SW-846 6010 D (Aqueous)

#### COPPER

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)

#### IRON

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
SW-846 6010 D (Aqueous)

#### LEAD

SW-846 7000 B (Aqueous)  
SW-846 7000 B (Non-Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)  
EPA 200.9, Rev. 2.2, 1994 (Aqueous)

#### MANGANESE

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
SW-846 6010 D (Aqueous)

#### MERCURY

EPA 245.1, Rev. 3.0, 1994 (Aqueous)  
SW-846 7470 A (Aqueous)  
SW-846 7471 B (Non-Aqueous)

#### NICKEL

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)

#### SELENIUM

EPA 200.7, Rev. 4.4, 1994 (Aqueous)  
EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)  
SW-846 6010 D (Aqueous)  
SW-846 6010 D (Non-Aqueous)

# North Carolina Wastewater/Groundwater Laboratory Certification

## Certified Parameters Listing

Lab Name: Schneider Laboratories Global Inc.  
Address: 2512 W. Cary Street  
Richmond, VA 23220-

Certificate Number: 593  
Effective Date: 1/1/2020  
Expiration Date: 12/31/2020  
Date of Last Amendment: 10/4/2018

The above named laboratory, having duly met the requirements of 15A NCAC 2H.0800, is hereby certified for the measurement of the parameters listed below.

### CERTIFIED PARAMETERS

EPA 200.9, Rev. 2.2, 1994 (Aqueous)

#### SILVER

EPA 200.7, Rev. 4.4, 1994 (Aqueous)

EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)

SW-846 6010 D (Aqueous)

SW-846 6010 D (Non-Aqueous)

#### THALLIUM

EPA 200.7, Rev. 4.4, 1994 (Aqueous)

SW-846 6010 D (Aqueous)

EPA 200.9, Rev. 2.2, 1994 (Aqueous)

#### VANADIUM

EPA 200.7, Rev. 4.4, 1994 (Aqueous)

SW-846 6010 D (Aqueous)

#### ZINC

EPA 200.7, Rev. 4.4, 1994 (Aqueous)

EPA 200.7, Rev. 4.4, 1994 (Non-Aqueous)

SW-846 6010 D (Aqueous)

**ROY COOPER**  
*Governor*  
**MICHAEL S. REGAN**  
*Secretary*  
**LINDA CULPEPPER**  
*Interim Director*



**MEMORANDUM**

DATE: December 6, 2019  
TO: Certified Laboratories  
FROM: Dana Satterwhite, Laboratory Certification Branch Program Supervisor  
SUBJECT: 2020 North Carolina Wastewater/Groundwater Laboratory Certification Renewal

The Department of Environmental Quality, in accordance with the provisions of NC G.S. 143-215.3 (a) (10), 15 NCAC 2H .0800, is pleased to renew Certification for your laboratory to perform specified environmental analyses.

Enclosed is an updated Certificate and Certified Parameters Listing (CPL) describing the scope of your accreditation. **Please review the CPL carefully to ensure that your laboratory is certified for all parameter methods required to properly meet your Certification needs. It is also recommended that you verify the scope of accreditation of any contract or subcontract labs at least annually to ensure they maintain certification with our program for the appropriate parameters.**

Contact us at 919-733-3908 if you have questions or need additional information.

Attachment

cc: Master File

